

測定月報

測定日: 2016/10

(1/5)

日	モニタ											
	A001 正門警備所 (DM74)			A002 正門警備所 (DM44)			A003 免震重要棟(DM-53)			A004 免震重要棟(DM-57)		
	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)
1	5.84E-06	1.0	100	4.85E-06	1.2	100	4.21E-06	9.7	100	3.93E-06	7.6	100
2	6.94E-06	1.0	100	5.81E-06	1.2	100	5.13E-06	9.7	100	4.82E-06	7.7	100
3	9.25E-06	1.1	100	8.13E-06	1.1	100	6.37E-06	9.6	100	6.04E-06	7.8	100
4	5.16E-06	1.0	100	4.28E-06	1.1	100	4.04E-06	9.9	100	3.87E-06	7.7	100
5	5.95E-06	1.1	100	4.97E-06	1.1	100	4.79E-06	9.5	100	4.40E-06	7.7	100
6	3.22E-06	1.0	100	2.72E-06	1.1	100	2.72E-06	9.6	100	2.54E-06	7.6	100
7	5.85E-06	1.1	100	4.93E-06	1.2	100	4.37E-06	9.8	100	4.36E-06	7.5	100
8	3.74E-06	0.9	100	2.97E-06	1.3	100	2.99E-06	9.6	100	2.79E-06	7.7	100
9	4.77E-06	1.0	100	4.02E-06	1.1	100	3.87E-06	9.7	100	3.71E-06	7.6	100
10	5.67E-06	1.1	100	4.80E-06	1.2	100	4.61E-06	9.7	100	4.17E-06	7.8	100
11	6.61E-06	1.0	100	5.50E-06	1.2	100	5.30E-06	9.6	100	4.87E-06	7.8	100
12	6.95E-06	1.0	100	5.82E-06	1.2	100	5.48E-06	9.6	100	4.90E-06	7.9	100
13	7.02E-06	1.0	100	5.63E-06	1.2	100	5.12E-06	9.7	100	4.68E-06	7.8	100
14	5.89E-06	1.0	100	*1 4.87E-06	1.2	100	4.58E-06	9.8	100	4.27E-06	7.8	100
15	7.51E-06	1.1	100	6.21E-06	1.1	100	5.80E-06	9.8	100	5.43E-06	7.7	100
16	6.96E-06	1.0	100	5.67E-06	1.2	100	5.06E-06	9.8	100	4.79E-06	7.6	100
17	4.04E-06	1.0	100	3.36E-06	1.2	100	2.74E-06	9.8	100	2.73E-06	7.5	100
18	6.29E-06	0.9	100	5.02E-06	1.2	100	4.64E-06	9.5	100	4.43E-06	7.5	100
19	8.36E-06	1.0	100	6.84E-06	1.1	100	5.86E-06	9.8	100	5.76E-06	7.7	100
20	1.01E-05	1.0	100	7.98E-06	1.2	100	7.19E-06	9.7	100	7.02E-06	7.5	100
21	*1 4.15E-06	1.1	100	3.37E-06	1.2	100	*1 3.29E-06	9.8	100	3.16E-06	7.7	100
22	7.56E-06	1.0	100	5.94E-06	1.2	100	5.44E-06	9.7	100	5.12E-06	7.8	100
23	5.13E-06	1.1	100	4.07E-06	1.2	100	3.99E-06	10.0	100	3.97E-06	7.6	100
24	4.69E-06	1.1	100	3.76E-06	1.2	100	3.93E-06	9.8	100	*1 3.76E-06	7.6	100
25	4.58E-06	1.0	100	3.51E-06	1.2	100	3.70E-06	9.6	100	3.43E-06	7.6	100
26	4.64E-06	1.0	100	3.64E-06	1.1	100	3.56E-06	9.6	100	3.35E-06	7.6	100
27	5.59E-06	1.1	100	4.63E-06	1.2	100	4.50E-06	9.8	100	4.39E-06	7.5	100
28	5.99E-06	1.1	100	4.66E-06	1.2	100	4.72E-06	9.8	100	4.54E-06	7.6	100
29	4.15E-06	1.0	100	3.21E-06	1.2	100	3.18E-06	9.7	100	3.05E-06	7.7	100
30	8.31E-06	1.0	100	6.28E-06	1.2	100	6.08E-06	9.7	100	5.82E-06	7.7	100
31	6.90E-06	1.0	100	5.31E-06	1.1	100	5.04E-06	9.8	100	4.84E-06	7.6	100
1ヶ月値	6.06E-06	1.0	100	4.93E-06	1.2	100	4.59E-06	9.7	100	4.35E-06	7.7	100
最大値	2.85E-05	1.1	102	2.86E-05	1.3	101	4.12E-05	10.0	102	2.60E-05	7.9	102
最大値日時	10/03 07:20	10/03 00:12	10/01 05:59	10/03 04:09	10/08 03:57	10/01 00:13	10/28 00:27	10/23 00:27	10/01 00:07	10/22 04:12	10/10 04:12	10/13 09:04
最小値	7.24E-07	0.9	98	0.00E+00	1.1	98	0.00E+00	9.5	98	0.00E+00	7.5	98
最小値日時	10/21 15:03	10/08 00:12	10/01 04:16	10/21 03:57	10/03 03:57	10/02 03:02	10/01 00:27	10/05 00:27	10/01 00:56	10/01 04:12	10/07 04:12	10/01 16:10

*1:ろ紙交換

*2:最大・最小値に関しては、1日1回BG測定による、影響があるため、値に関しては、参考値扱いとする

測定月報

測定日: 2016/10

(2/5)

日	モニタ											
	A005			A006			A007			A008		
	5/6号警備所前(DM55)			5/6号警備所前(DM59)			1号北側法面上(旧厚生棟)(DM-51)			1号北側法面上(旧厚生棟)(DM-47)		
	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)
1	3.95E-06	1.8	100	4.77E-06	1.4	100	5.10E-06	12.8	100	5.17E-06	9.9	100
2	5.08E-06	1.7	100	6.06E-06	1.5	100	6.24E-06	12.6	100	6.35E-06	9.8	100
3	6.27E-06	1.7	100	7.53E-06	1.6	100	7.72E-06	12.7	100	7.98E-06	9.7	100
4	4.10E-06	1.8	100	4.63E-06	1.5	100	5.08E-06	12.8	100	5.19E-06	9.6	100
5	4.74E-06	1.7	100	5.23E-06	1.7	100	5.92E-06	11.8	100	5.72E-06	9.8	100
6	2.63E-06	1.6	100	2.54E-06	1.9	100	3.36E-06	10.1	100	3.24E-06	9.7	100
7	4.29E-06	1.6	100	4.16E-06	2.0	100	*1 5.49E-06	10.3	100	*1 5.50E-06	9.4	100
8	3.08E-06	1.7	100	2.92E-06	2.1	100	3.97E-06	10.4	100	3.98E-06	9.6	100
9	3.98E-06	1.7	100	3.71E-06	2.1	100	4.99E-06	10.1	100	4.54E-06	9.7	100
10	4.25E-06	1.8	100	4.10E-06	2.0	100	5.32E-06	10.4	100	5.16E-06	9.6	100
11	5.05E-06	1.7	100	4.87E-06	1.9	100	6.45E-06	10.2	100	5.96E-06	9.5	100
12	5.30E-06	1.7	100	5.07E-06	2.0	100	6.78E-06	10.1	100	6.02E-06	9.8	100
13	4.94E-06	1.7	100	4.60E-06	2.0	100	6.46E-06	10.1	100	4.84E-06	10.5	100
14	4.33E-06	1.7	100	*1 4.17E-06	2.0	100	5.75E-06	10.1	100	4.38E-06	10.4	100
15	5.82E-06	1.7	100	5.52E-06	2.0	100	7.08E-06	10.4	100	5.85E-06	10.1	100
16	5.13E-06	1.7	100	4.91E-06	2.0	100	6.25E-06	10.3	100	4.96E-06	10.4	100
17	3.26E-06	1.8	100	3.20E-06	2.1	100	4.09E-06	10.5	100	3.63E-06	10.1	100
18	4.59E-06	1.8	100	4.41E-06	2.1	100	6.00E-06	10.0	100	4.70E-06	10.2	100
19	5.76E-06	1.7	100	5.55E-06	2.0	100	7.58E-06	10.2	100	5.95E-06	10.2	100
20	7.65E-06	1.8	100	7.27E-06	2.0	100	9.22E-06	10.3	100	7.29E-06	10.2	100
21	3.27E-06	1.6	100	3.05E-06	2.0	100	4.06E-06	10.2	100	3.26E-06	10.3	100
22	5.24E-06	1.7	100	4.93E-06	2.0	100	6.98E-06	10.0	100	5.40E-06	10.2	100
23	3.94E-06	1.7	100	3.79E-06	1.9	100	5.08E-06	10.2	100	4.03E-06	10.2	100
24	*1 3.62E-06	1.7	100	3.38E-06	2.1	100	4.87E-06	10.0	100	*1 3.80E-06	10.4	100
25	3.41E-06	1.8	100	3.34E-06	2.0	100	4.37E-06	11.0	100	3.69E-06	10.2	100
26	3.28E-06	1.8	100	3.24E-06	2.0	100	3.12E-06	12.4	100	3.35E-06	10.4	100
27	4.47E-06	1.7	100	*1 4.40E-06	2.0	100	*1 4.44E-06	11.8	100	4.41E-06	10.2	100
28	4.48E-06	1.7	100	4.32E-06	2.0	100	4.32E-06	12.0	100	4.31E-06	10.4	100
29	3.16E-06	1.6	100	3.06E-06	1.8	100	3.00E-06	12.0	100	3.07E-06	10.2	100
30	5.84E-06	1.7	100	5.70E-06	1.9	100	5.52E-06	12.3	100	5.74E-06	10.3	100
31	4.82E-06	1.8	100	4.79E-06	1.9	100	4.87E-06	12.0	100	5.08E-06	10.1	100
1ヶ月値	4.51E-06	1.7	100	4.49E-06	1.9	100	5.47E-06	11.0	100	4.92E-06	10.0	100
最大値	2.10E-05	1.8	102	2.58E-05	2.1	103	4.94E-05	12.8	102	4.22E-05	10.6	102
最大値日時	10/03 06:46	10/01 00:42	10/01 04:22	10/03 06:50	10/08 04:27	10/02 06:48	10/05 15:07	10/01 00:57	10/01 01:50	10/24 13:59	10/13 04:42	10/01 03:59
最小値	0.00E+00	1.6	98	0.00E+00	1.4	96	0.00E+00	10.0	97	0.00E+00	9.4	97
最小値日時	10/01 00:42	10/01 00:01	10/01 01:10	10/05 14:20	10/01 04:27	10/12 09:20	10/01 00:57	10/18 00:57	10/28 11:38	10/03 04:42	10/07 04:42	10/11 13:08

*1: ろ紙交換

*2: 最大・最小値に関しては、1日1回BG測定による、影響があるため、値に関しては、参考値扱いとする

測定月報

測定日: 2016/10

(3/5)

日	モニタ											
	A009			A010			A011			A012		
	1/2号西側法面上 (DM75)			1/2号西側法面上 (DM65)			3/4号西側法面上 (DM45)			3/4号西側法面上 (DM54)		
	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)
1	5.84E-06	7.4	100	4.28E-06	8.0	100	4.41E-06	5.3	100	4.22E-06	6.2	100
2	6.83E-06	7.3	100	5.25E-06	7.9	100	5.22E-06	5.4	100	5.25E-06	6.0	100
3	8.48E-06	7.3	100	6.50E-06	7.8	100	6.56E-06	5.4	100	6.75E-06	6.1	100
4	5.62E-06	7.4	100	4.18E-06	8.0	100	4.06E-06	5.6	100	3.83E-06	6.4	100
5	6.11E-06	7.4	100	4.72E-06	7.8	100	4.49E-06	5.5	100	4.55E-06	6.1	100
6	3.51E-06	7.3	100	2.69E-06	7.9	100	2.57E-06	5.4	100	2.53E-06	6.2	100
7	5.97E-06	7.4	100	4.42E-06	8.1	100	4.46E-06	5.2	100	4.44E-06	6.0	100
8	3.97E-06	7.4	100	2.96E-06	8.1	100	3.03E-06	5.2	100	3.02E-06	6.0	100
9	5.23E-06	7.2	100	3.88E-06	7.9	100	3.85E-06	5.3	100	3.85E-06	6.1	100
10	5.91E-06	7.3	100	4.46E-06	8.0	100	4.30E-06	5.2	100	4.42E-06	5.9	100
11	6.85E-06	7.4	100	5.09E-06	8.0	100	5.04E-06	5.2	100	5.21E-06	5.8	100
12	7.21E-06	7.2	100	5.48E-06	7.8	100	5.28E-06	5.2	100	5.45E-06	5.8	100
13	6.69E-06	7.4	100	5.06E-06	8.0	100	5.06E-06	5.0	100	5.24E-06	5.6	100
14	6.23E-06	7.1	100	4.56E-06	8.1	100	*1 4.60E-06	5.1	100	4.47E-06	5.8	100
15	7.65E-06	7.4	100	5.65E-06	8.0	100	5.69E-06	5.0	100	5.83E-06	5.8	100
16	6.56E-06	7.4	100	4.97E-06	7.9	100	5.09E-06	5.0	100	5.17E-06	5.9	100
17	4.03E-06	7.2	100	2.94E-06	7.8	100	2.89E-06	5.1	100	2.92E-06	5.8	100
18	6.21E-06	7.3	100	4.68E-06	8.0	100	4.62E-06	5.1	100	4.41E-06	6.0	100
19	8.31E-06	7.3	100	6.20E-06	7.9	100	6.08E-06	5.1	100	6.13E-06	5.9	100
20	1.02E-05	7.2	100	7.43E-06	7.9	100	7.62E-06	5.1	100	7.66E-06	6.0	100
21	4.38E-06	7.4	100	3.38E-06	7.9	100	3.19E-06	5.2	100	3.31E-06	5.8	100
22	7.18E-06	7.4	100	5.44E-06	8.0	100	5.24E-06	5.1	100	5.58E-06	5.7	100
23	5.59E-06	7.3	100	4.25E-06	7.8	100	4.05E-06	5.2	100	4.15E-06	6.0	100
24	*1 5.06E-06	7.4	100	*1 3.87E-06	7.8	100	3.62E-06	5.2	100	*1 3.82E-06	5.9	100
25	4.82E-06	7.4	100	3.55E-06	8.1	100	3.53E-06	5.0	100	3.53E-06	5.9	100
26	4.91E-06	7.3	100	3.72E-06	7.8	100	3.50E-06	5.2	100	3.45E-06	6.0	100
27	6.07E-06	7.4	100	4.76E-06	7.8	100	4.37E-06	5.1	100	4.45E-06	5.8	100
28	6.46E-06	7.2	100	4.72E-06	7.9	100	4.36E-06	4.9	100	4.43E-06	5.6	100
29	4.52E-06	7.3	100	3.39E-06	7.8	100	3.23E-06	4.9	100	3.19E-06	5.6	100
30	8.65E-06	7.2	100	6.13E-06	7.9	100	6.13E-06	4.8	100	5.95E-06	5.6	100
31	6.97E-06	7.3	100	4.99E-06	7.9	100	5.05E-06	4.7	100	4.92E-06	5.7	100
1ヶ月値	6.19E-06	7.3	100	4.63E-06	7.9	100	4.56E-06	5.2	100	4.58E-06	5.9	100
最大値	2.83E-05	7.5	102	2.87E-05	8.2	102	2.20E-05	5.6	102	2.42E-05	6.5	102
最大値日時	10/03 07:05	10/24 01:12	10/01 00:09	10/03 04:57	10/25 04:57	10/01 08:40	10/03 07:19	10/04 01:27	10/02 18:07	10/03 05:43	10/04 05:12	10/02 08:39
最小値	0.00E+00	7.1	97	0.00E+00	7.7	98	0.00E+00	4.7	98	0.00E+00	5.6	98
最小値日時	10/01 01:13	10/14 01:12	10/01 04:24	10/01 04:57	10/26 04:57	10/01 09:55	10/05 01:27	10/31 01:27	10/08 01:38	10/01 05:12	10/13 05:12	10/01 01:52

*1:ろ紙交換

*2:最大・最小値に関しては、1日1回BG測定による、影響があるため、値に関しては、参考値扱いとする

測定月報

測定日: 2016/10

(4/5)

日	モニタ											
	A013			A014			A015			A016		
	4号南側法面上[3号南側](DM69)			4号南側法面上[3号南側](DM56)			登録S休憩所 (DM52)			登録S休憩所 (DM68)		
	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)
1	4.38E-06	1.5	100	4.09E-06	1.2	100	4.51E-06	2.5	100	6.03E-06	2.4	100
2	5.43E-06	1.6	100	5.07E-06	1.2	100	5.45E-06	2.6	100	7.34E-06	2.4	100
3	6.96E-06	1.6	100	6.46E-06	1.2	100	7.51E-06	2.5	100	9.97E-06	2.5	100
4	4.28E-06	1.6	100	3.98E-06	1.2	100	3.87E-06	2.6	100	5.10E-06	2.4	100
5	4.75E-06	1.6	100	4.41E-06	1.3	100	4.54E-06	2.6	100	6.06E-06	2.4	100
6	2.69E-06	1.6	100	2.56E-06	1.2	100	2.38E-06	2.6	100	3.27E-06	2.4	100
7	4.70E-06	1.5	100	4.49E-06	1.1	100	4.42E-06	2.6	100	6.11E-06	2.4	100
8	2.96E-06	1.6	100	2.87E-06	1.2	100	2.90E-06	2.5	100	3.88E-06	2.3	100
9	4.00E-06	1.5	100	3.85E-06	1.1	100	3.60E-06	2.6	100	4.82E-06	2.3	100
10	4.56E-06	1.5	100	4.36E-06	1.1	100	4.20E-06	2.6	100	5.86E-06	2.3	100
11	5.30E-06	1.5	100	5.05E-06	1.2	100	4.97E-06	2.5	100	6.64E-06	2.3	100
12	5.47E-06	1.5	100	5.16E-06	1.2	100	5.32E-06	2.6	100	7.29E-06	2.4	100
13	5.03E-06	1.5	100	4.84E-06	1.2	100	5.39E-06	2.4	100	7.16E-06	2.4	100
14	*1 4.58E-06	1.5	100	4.45E-06	1.1	100	4.45E-06	2.7	100	6.13E-06	2.4	100
15	6.00E-06	1.4	100	5.64E-06	1.2	100	5.97E-06	2.6	100	8.13E-06	2.4	100
16	5.23E-06	1.5	100	4.89E-06	1.2	100	5.73E-06	2.6	100	7.91E-06	2.4	100
17	2.79E-06	1.5	100	2.67E-06	1.2	100	3.15E-06	2.5	100	4.36E-06	2.4	100
18	4.64E-06	1.5	100	4.31E-06	1.2	100	4.78E-06	2.5	100	6.44E-06	2.4	100
19	6.36E-06	1.5	100	6.06E-06	1.1	100	6.15E-06	2.5	100	8.39E-06	2.4	100
20	7.45E-06	1.7	100	7.11E-06	1.3	100	7.53E-06	2.7	100	1.03E-05	2.5	100
21	3.29E-06	1.6	100	*1 3.26E-06	1.2	100	*1 3.16E-06	2.5	100	*1 4.32E-06	2.3	100
22	5.51E-06	1.6	100	5.25E-06	1.2	100	5.40E-06	2.6	100	7.44E-06	2.4	100
23	4.26E-06	1.6	100	4.03E-06	1.1	100	3.93E-06	2.5	100	5.16E-06	2.4	100
24	3.88E-06	1.5	100	3.80E-06	1.1	100	3.61E-06	2.5	100	4.90E-06	2.4	100
25	3.53E-06	1.5	100	3.41E-06	1.2	100	3.38E-06	2.5	100	4.74E-06	2.3	100
26	3.51E-06	1.6	100	3.38E-06	1.1	100	3.45E-06	2.4	100	4.61E-06	2.4	100
27	4.52E-06	1.6	100	4.40E-06	1.2	100	4.52E-06	2.4	100	5.91E-06	2.4	100
28	4.58E-06	1.5	100	4.44E-06	1.2	100	4.39E-06	2.6	100	5.97E-06	2.4	100
29	3.31E-06	1.5	100	3.20E-06	1.1	100	3.02E-06	2.6	100	4.06E-06	2.5	100
30	6.11E-06	1.5	100	5.90E-06	1.1	100	6.14E-06	2.6	100	8.35E-06	2.3	100
31	5.14E-06	1.5	100	4.85E-06	1.2	100	5.20E-06	2.5	100	6.88E-06	2.4	100
1ヶ月値	4.68E-06	1.5	100	4.46E-06	1.2	100	4.61E-06	2.5	100	6.25E-06	2.4	100
最大値	2.27E-05	1.7	102	2.35E-05	1.3	102	2.32E-05	2.7	102	4.05E-05	2.5	102
最大値日時	10/03 06:55	10/20 01:42	10/01 10:45	10/03 06:09	10/05 05:27	10/01 01:10	10/03 06:58	10/14 01:57	10/02 08:05	10/03 06:05	10/03 05:42	10/24 21:21
最小値	0.00E+00	1.4	98	0.00E+00	1.1	98	0.00E+00	2.4	98	0.00E+00	2.2	98
最小値日時	10/02 01:42	10/15 01:42	10/01 02:05	10/05 05:27	10/01 00:01	10/01 08:56	10/02 01:57	10/13 01:57	10/02 18:16	10/01 05:42	10/21 14:48	10/10 08:14

*1:ろ紙交換

*2:最大・最小値に関しては、1日1回BG測定による、影響があるため、値に関しては、参考値扱いとする

測定月報

測定日: 2016/10

(5/5)

日	モニタ											
	A019 シールド中央制御室(DM-49)			A020 シールド中央制御室(DM-60)			A021 Hタンクエリア (DM67)			A022 Hタンクエリア (DM64)		
	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)	濃度 (Bq/cm ³)	計数率 (cps)	流量 (L/min)
1	5.34E-06	3.1	100	5.09E-06	3.4	100	6.15E-06	3.5	100	4.62E-06	3.1	100
2	6.53E-06	3.1	100	6.35E-06	3.3	100	7.30E-06	3.4	100	5.36E-06	3.2	100
3	8.50E-06	3.1	100	8.24E-06	3.4	100	9.66E-06	3.4	100	7.24E-06	3.2	100
4	5.20E-06	3.0	100	4.91E-06	3.3	100	5.39E-06	3.5	100	4.12E-06	3.2	100
5	5.65E-06	3.3	100	5.54E-06	3.4	100	6.39E-06	3.5	100	4.88E-06	3.1	100
6	3.22E-06	3.2	100	3.11E-06	3.3	100	3.54E-06	3.4	100	2.63E-06	3.2	100
7	5.63E-06	3.2	100	5.55E-06	3.3	100	6.18E-06	3.4	100	4.59E-06	3.2	100
8	3.55E-06	3.2	100	3.37E-06	3.4	100	3.93E-06	3.5	100	2.92E-06	3.3	100
9	4.74E-06	3.1	100	4.42E-06	3.4	100	5.11E-06	3.4	100	3.81E-06	3.2	100
10	5.51E-06	3.2	100	5.37E-06	3.4	100	6.08E-06	3.3	100	4.61E-06	3.0	100
11	*1 6.49E-06	3.2	100	*1 6.22E-06	3.4	100	7.09E-06	3.3	100	5.24E-06	3.2	100
12	6.79E-06	3.3	100	6.29E-06	3.4	100	7.35E-06	3.5	100	5.63E-06	3.2	100
13	6.46E-06	3.2	100	5.94E-06	3.4	100	7.22E-06	3.3	100	5.45E-06	3.1	100
14	5.66E-06	3.3	100	5.26E-06	3.4	100	6.48E-06	3.3	100	4.71E-06	3.2	100
15	7.40E-06	3.1	100	6.75E-06	3.4	100	7.89E-06	3.4	100	5.94E-06	3.2	100
16	6.41E-06	3.2	100	6.06E-06	3.3	100	6.86E-06	3.5	100	5.25E-06	3.3	100
17	3.54E-06	3.1	100	3.15E-06	3.4	100	3.95E-06	3.5	100	2.90E-06	3.2	100
18	4.98E-06	2.9	100	5.32E-06	3.4	100	6.40E-06	3.4	100	4.87E-06	3.1	100
19	5.46E-06	2.7	100	7.42E-06	3.3	100	8.38E-06	3.4	100	6.38E-06	3.1	100
20	6.60E-06	2.7	100	8.69E-06	3.4	100	1.01E-05	3.4	100	7.54E-06	3.3	100
21	2.97E-06	2.7	100	3.82E-06	3.5	100	*1 4.51E-06	3.5	100	*1 3.47E-06	3.2	100
22	5.05E-06	2.5	100	6.54E-06	3.3	100	7.51E-06	3.5	100	5.74E-06	3.2	100
23	3.64E-06	2.8	100	4.88E-06	3.3	100	5.53E-06	3.5	100	4.11E-06	3.3	100
24	3.35E-06	2.6	100	4.41E-06	3.4	100	5.15E-06	3.3	100	3.99E-06	3.1	100
25	3.02E-06	2.7	100	4.05E-06	3.3	100	4.72E-06	3.6	100	3.60E-06	3.3	100
26	2.89E-06	2.8	100	3.44E-06	3.1	100	4.73E-06	3.5	100	3.56E-06	3.2	100
27	*1 3.88E-06	2.7	100	4.38E-06	3.2	100	6.21E-06	3.5	100	4.68E-06	3.2	100
28	3.94E-06	2.7	100	4.44E-06	3.2	100	6.21E-06	3.5	100	4.71E-06	3.2	100
29	2.80E-06	2.7	100	3.24E-06	3.1	100	4.40E-06	3.5	100	3.26E-06	3.2	100
30	5.28E-06	2.7	100	5.90E-06	3.1	100	8.49E-06	3.5	100	6.41E-06	3.1	100
31	4.25E-06	2.8	100	4.84E-06	3.2	100	6.95E-06	3.5	100	5.26E-06	3.2	100
1ヶ月値	4.99E-06	3.0	100	5.26E-06	3.3	100	6.32E-06	3.4	100	4.76E-06	3.2	100
最大値	2.74E-05	3.3	103	3.94E-05	3.5	103	3.23E-05	3.6	102	3.18E-05	3.4	102
最大値日時	10/03 07:13	10/05 02:27	10/21 22:25	10/03 06:30	10/08 06:12	10/17 04:32	10/03 02:53	10/21 02:42	10/14 04:52	10/03 06:49	10/25 06:27	10/04 02:45
最小値	0.00E+00	2.5	97	0.00E+00	3.0	97	0.00E+00	3.3	98	0.00E+00	3.0	98
最小値日時	10/02 02:27	10/22 02:27	10/19 00:46	10/06 06:12	10/29 06:12	10/04 03:45	10/01 02:42	10/01 00:01	10/07 18:35	10/01 06:27	10/10 06:27	10/04 05:46

*1: ろ紙交換

*2: 最大・最小値に関しては、1日1回BG測定による、影響があるため、値に関しては、参考値扱いとする